



©2025 ECD

THIS DRAWING OR DOCUMENT IS THE PROPERTY OF ECD. AND EMBODIES CONFIDENTIAL PROPRIETARY DESIGN OR DATA IN WHICH ECD RETAINS ALL PATENT AND OTHER RIGHTS INCLUDING EXCLUSIVE RIGHTS OF USE, MANUFACTURE OR SALE. IT IS SUBMITTED UNDER A CONFIDENTIAL RELATIONSHIP FOR A SPECIFIC PURPOSE AND THE RECIPIENT AGREES BY ACCEPTING THIS DRAWING OR DOCUMENT NOT TO SUPPLY OR DISCLOSE ANY INFORMATION REGARDING IT TO ANY OTHER PARTY OR TO USE IT FOR ANY OTHER PURPOSE THAN THE SPECIFIC PURPOSE FOR WHICH SUBMITTED

ALL COMPONENTS, MATERIALS AND MANUFACTURING PRACTICES MUST MEET THE CURRENT RESTRICTION OF THE USE OF CERTAIN HAZARDOUS SUBSTANCES IN ELECTRICAL AND ELECTRONIC EQUIPMENT (ROHS) DIRECTIVE OF THE EUROPEAN PARLIAMENT

## CALIBRATION INFORMATION

Because the M.O.L.E.<sup>®</sup> Thermal Profiler is made with precision components with high temperature stability and tight tolerances, the analog-to-digital converters remains stable for years. High quality components together with software algorithms based on the **ITS-90\*** standard for **Type K** thermocouples have been provided to yield the specified accuracy and long-term stability. Each unit has been tested at the factory before it is shipped.

Good thermal quality programs require periodic calibration to show the Thermal Profiler continues to remain in calibration using a temperature standard.



Do not attempt to calibrate the V-M.O.L.E.<sup>™</sup> 2 Thermal Profiler if you have never used a thermocouple simulator, or you are unsure of the accuracy of your Thermocouple Simulator. (Contact ECD for further assistance).

---

\* ITS-90 - International Temperature Scale of 1990

## DEFINITIONS:

- **Calibration:** The name (noun) of the procedure that determines the accuracy of a measurement device by comparing it to a known standard.
- **Calibrate:** To perform (verb) the procedure of calibration, which involves measuring against a standard. It does not include adjustment, though adjustment may follow calibration if needed.
- **Adjustment (Calibration Adjustment):** The procedure of modifying the device to bring its measurements closer to the standard.
- **Certificate of Calibration (COC):** The resulting documentation issued by a laboratory that states the device's accuracy, the instruments used and conditions while the procedure was performed.
  - Accredited COC is Issued by a laboratory that follows standard practices has been assessed and accredited by an independent accreditation body.
  - Unaccredited COC is issued by a laboratory that may follow standard practices but has not been formally accredited by in independent body.

## EQUIPMENT REQUIRED:

1. A Thermocouple Standard which may be either:
  - Thermocouple Simulator\* with accuracy  $\pm 0.1^{\circ}\text{C}$  and output impedance  $\leq 0.1\Omega$
  - Voltage Reference with accuracy  $\pm 3\mu\text{V}$  and output impedance  $\leq 0.1\Omega$ , Ice-point Cell with accuracy  $\pm 1^{\circ}\text{C}$  or better and TRP
2. Thermocouple harness (special limits of error)
  - ECD Item #E65-0509-26
3. Foam insulator (to wrap around the M.O.L.E. Thermal Profiler thermocouple connections and provide access to the ON/OFF and Record buttons)
4. [M.O.L.E. MAP Software](#) (installed on PC)
5. Optional – [MAP Machine and KPI Template files](#)
  - MOLECalibration.xmr
  - MOLECalibrationTemplate.tpf


---

\* ITS-90 - International Temperature Scale of 1990


## SETUP:

The equipment you use for the calibration determines the setup procedure.

### Voltage Reference, Ice-point Cell and TRP connection:



### Thermocouple Simulator connection:



## PROCEDURE: CALIBRATION VERIFICATION

1. Connect the V-M.O.L.E.™ 2 Thermal Profiler to the thermocouple harness ECD Item #E65-0509-26.
2. Wrap the thermocouple connection of the M.O.L.E. with foam insulator.
3. Turn ON the M.O.L.E. by pressing the **ON/OFF** button.




4. Start M.O.L.E. MAP and set M.O.L.E. recording parameters as follows:
  - Log Interval: 1 sec
  - All channels ON and type K
5. Set the Thermocouple Standard to -100°C (-3.554mV). **Allow at least 1 minute for thermal stabilization.**



Prior to continuing it is helpful to confirm if any channels are reading open (----.-) and temperatures are reading near -100°C (-3.554mV) using the Instrument Status command in M.O.L.E. MAP software.

6. Start M.O.L.E. recording, and allow to log for at least 30 seconds at -100°C or user determined value.
7. Change the Thermocouple Standard to 100°C (4.096mV) or user determined value, and allow it to continue logging for at least 30 more seconds.
8. Change the Thermocouple Standard to 250°C (10.153mV) or user determined value, and allow it to continue logging for at least 30 more seconds.
9. Change the Thermocouple Standard to 1000°C (41.276mV) or user determined value, and allow it to continue logging for at least 30 more seconds.
10. Stop M.O.L.E. recording.

11. Download the data run from the M.O.L.E. to the M.O.L.E. MAP Software.



12. Confirm the M.O.L.E. recorded the temperatures  $\pm 1^{\circ}\text{C}$  from the Thermocouple Standard (guard-band using your standard's accuracy, and or uncertainty as you prefer).

- If within spec, Calibration Verification complete.
- If out of spec, Perform Calibration Adjustment procedure.


## PROCEDURE: CALIBRATION ADJUSTMENT

1. Connect the V-M.O.L.E.™ 2 Thermal Profiler to the thermocouple harness ECD Item #E65-0509-26.
2. Wrap the thermocouple connection of the M.O.L.E. with foam insulator.
3. Turn ON the M.O.L.E. by pressing the **ON/OFF** button.




4. Set the Thermocouple Standard to 0°C (0.000mV). Allow at least 1 minute for thermal stabilization.

5. Locate the M.O.L.E. MAP software desktop shortcut and create a copy. Add the following to the Target: “-calibrate” which includes a space after the ...exe






6. Start the M.O.L.E. MAP Software and select **Setup Instrument** from the **M.O.L.E.** menu.
7. Press the **Start** button to begin.



8. Follow the instructions on each dialog when prompted.



9. Success or failure will be prompted at the end of each step.
  - If successful Calibration Adjustment complete. Repeat the Calibration Verification to confirm.
  - If not, confirm again that ALL channels are reading at or near the current standard's setting, checking for bad connections, and repeat the Calibration Adjustment procedure.